



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description: BIOPHEN UFH CONTROL PLASMA
Product Code: 223101
Lot No.: FA23281A
Expiration date: 2024-02-12

| Components | Quantity | Lot | Expiration Date |
|------------|----------|------------|-----------------|
| C1 | 6 vials | FA1495BX19 | 2024-02-12 |
| C2 | 6 vials | FA1754BY19 | 2024-03-11 |

| Analytical Data | Specification | Result |
|-----------------|--|--------------------------|
| Concentration | C1 : 0.10 -0.30 IU/mL C2 : 0.40 -0.60 IU/mL | 0.21 IU/mL 0.49 IU/mL |

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2021-12-13

Quality Control Manager: S.LECOURT