



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN ARIXTRA CONTROL PLASMA
Product Code : 224001
Lot No. : FA165819
Expiration date : 2024-02-26

Components	Quantity	Lot	Expiration Date
C1	6 vials	FA1658CR19	2024-02-26
C2	6 vials	FA1658CS19	2024-02-26

Analytical Data	Specification	Result
Concentration	C1 : 0.30 -0.50 µg/ml C2: 1.05 -1.35 µg/ml	0.41 µg/ml 1.20 µg/ml

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2021-09-07

Quality Control Manager :

S.LECOURT