



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN UFH CONTROL PLASMA
Product Code : 223101
Lot No. : FA1872
Expiration date : 2024-03-11

Components	Quantity	Lot	Expiration Date
C1	6 vials	FA1755BX1A	2024-03-11
C2	6 vials	FA1754BY19	2024-03-11

Analytical Data	Specification	Result
Concentration	C1 : 0.10 -0.30 IU/ml C2 : 0.40 -0.60 IU/ml	0.19 IU/mL 0.49 IU/mL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2021-10-21

Quality Control Manager : S.LECOURT