



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description: BIOPHEN UFH CONTROL PLASMA
Product Code: 223101
Lot No.: FB1433
Expiration date: 2024-03-11

Components	Quantity	Lot	Expiration Date
C1	6 vials	FA1755BX1A	2024-03-11
C2	6 vials	FB0723BY1B	2024-10-14

Analytical Data	Specification	Result
Concentration	C1: 0.10 -0.30 IU/mL C2: 0.40 -0.60 IU/mL	0.19 IU/mL 0.53 IU/mL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2022-08-29

Quality Control Manager: S. LECOURT