



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description: CEPHEN 2.5
Product Code: CK512K
Lot No.: FB0576A1
Expiration date: 2024-03-31

Components	Quantity	Lot	Expiration Date
CEPHEN 2.5	6 vials	FB0576KBA1	2024-03-31

Analytical Data	Specification	Result
aPTT on Normal Plasma	Ratio < 1.20	1.00 (STAR) 1.00 (CS)

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2022-05-02

Quality Control Manager: S. LECOURT