



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : Prothrombin Deficient Plasma
Product Code : DP010A/K
Lot No. : FA149119
Expiration date : 2024-03-05

Components	Quantity	Lot	Expiration Date
Prothrombin Deficient Plasma	1 vial (DP010A) 6 vials (DP010K)	FA1491JO19	2024-03-05

Analytical Data	Specification	Result
Calibration curve	CT (0%) > 100 sec CT (100%) < 30 sec Thromboplastin : Neoplastin (Stago)	> 250 25.6
Concentration FII	< 1%	< 1%

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2021-09-27

Quality Control Manager : S.LECOURT