



HYPHEN BioMed  
 155 rue d'Eragny  
 95000 Neuville-sur-Oise  
 France

## CERTIFICATE OF ANALYSIS

**Product Description : ZYMUTEST Protein C**

**Product Code : RK027A**

**Lot No. : FA1736**

**Expiration date : 2024-03-04**

Components	Quantity	Lot	Expiration Date
Anti (h) Protein C pre-coated plate	12x8 wells	FA1913	2024-03-24
Protein C Standard	3 vials	FA1975	2024-04-08
Protein C Control I	1 vial	FA1977	2024-04-08
Protein C Control II	1 vial	FA1976	2024-04-08
Anti-(h)-Protein C HRP immunoconjugate	3 vials	FA2015	2024-04-15
Sample diluent	2x50 ml	FA1737	2025-03-10
Wash solution	1x50 ml	F2001234	2024-03-04
Conjugate diluent	1x25 ml	FA1121	2024-11-18
TMB	1x25 ml	210302D01	2025-03-30
Sulfuric Acid 0,45M	1x6 ml	FA0938	2024-10-14

Analytical Data	Specification	Result
Reactivity	A450 for 100% $\geq 1.40$	1.61
Calibration Curve and Accuracy	Controls Measured Values within acceptance ranges	PASS
Blank	$< 0.100$	0.010
Detection threshold	$\leq 5 \%$	$< 1$

### Calibration Curve for information

Conc.	C	C/2	C/4	C/10	C/20
OD	1.83	1.32	0.90	0.45	0.24

	Acceptance Range	MV
CI	[ 69 - 93 ]	83
CII	[ 39 - 59 ]	49

MV = Measured value

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval : 2021-11-16

Quality Control Manager : S.LECOURT

