



A Company of Sysmex Group



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description:** BIOPHEN a2 ANTIPLASMIN (LRT)  
**Product Code:** 220502  
**Lot No.:** FB07971E  
**Expiration date:** 2024-04-28

Components	Quantity	Lot	Expiration Date
R1 :Plasmin	3 vials	FB0797171E	2024-04-28
R2 : SPm-41 Substrate	3 vials	FB0797181E	2024-04-28

Analytical Data	Specification	Result
Calibration curve : reactivity, slope, linearity and accuracy on CS	Delta A405(100-50%)> 0.05 A405/min(100%) = $0.25 \pm 0.10$ $R^2 \geq 0.98$ Measured values for controls in acceptance ranges	0.09 0.27 1.00 PASS
Detection threshold (LOQ) on CS	$\leq 10\%$	3.5%

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval: 2022-05-24

Quality Control Manager: **S.LECOURT**