



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description: BIOPHEN AT (Anti IIa)
Product Code: 221122
Lot No.: FA206818
Expiration date: 2024-04-26

Components	Quantity	Lot	Expiration Date
R1 : Bovin Thrombin	2 vials	FA2068IT18	2024-05-13
R2 : Substrate	2 vials	FA2068IU18	2024-05-06
R3: Tris heparin Buffer	2 vials	FA2068IV18	2024-04-26

Analytical Data	Specification	Result
Calibration curve : reactivity, linearity, slope and accuracy	Delta A405 (0-100%) ≥ 0.80 $R^2 \geq 0.98$ Measured values for controls in acceptance ranges	1.315 0.994 PASS
Detection Threshold	$\leq 10\%$	$< 1\%$

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2021-12-07

Quality Control Manager: S. LECOURT