



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description: BIOPHEN Plasminogen LRT
Product Code: 221511
Lot No.: FA19501H
Expiration date: 2024-04-05

Components	Quantity	Lot	Expiration Date
R1 : Streptokinase	3 vials	FA1950b81H	2024-04-05
R2 : Substrate	3 vials	FA1950b91H	2024-04-05

Analytical Data	Specification	Result
Calibration curve : reactivity, slope, linearity and accuracy on CS	A405(100%) = 0.45 ± 0.20	0.409
	Delta A405 (100-50%) > 0.15	0.207
	R ² ≥ 0.98	1.000
	Measured values for controls in acceptance ranges	PASS
Detection Threshold	< 5 %	2%

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2021-10-28

Quality Control Manager: S. LECOURT