



A Company of Sysmex Group



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description :** BIOPHEN LMWH CONTROL LOW  
**Product Code :** 223701  
**Lot No. :** FA2148  
**Expiration date :** 2024-04-01

Components	Quantity	Lot	Expiration Date
CI	6 vials	FA1933BQ1D	2024-04-01
CII	6 vials	FA1934BR1B	2024-04-01

Analytical Data	Specification	Result
Concentration	CI : 0.15 -0.35 IU/ml CII : 0.35 -0.65 IU/ml	0.26 IU/mL 0.50 IU/mL

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval : 2021-11-16

Quality Control Manager : S.LECOURT