



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description :** ARACHIDONIC ACID  
**Product Code :** AG003K  
**Lot No. :** FB1736  
**Expiration date :** 2024-04-17

Components	Quantity	Lot	Expiration Date
ARACHIDONIC ACID	3 vials	FB1736	2024-04-17

Analytical Data	Specification	Result
Reactivity (% AgregMax) 1.0 mM on CS	≥ 50%	92%
Reactivity (% AgregMax) 1.0 mM on Agregometer	≥ 50%	83%

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval : 2022-11-29

Quality Control Manager : **S. LECOURT**

