



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description: BIOPHEN RIVAROXABAN CONTROL PLASMA
Product Code: 224501
Lot No.: FA22021A / FA220313
Expiration date: 2024-05-17

Components	Quantity	Lot	Expiration Date
C1	6 vials	FA2202BD1A / FA220313	2024-05-17
C2	6 vials	FA2202BE1A / FA220313	2024-05-17

Analytical Data	Specification	Result
Concentration	C1 : 50 -150 ng/mL C2 : 250 -350 ng/mL	102 325

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2021-12-13

Quality Control Manager: S. LECOURT