



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description: BIOPHEN™ Dabigatran Control Plasma
Product Code: 224701
Lot No.: FA22591D
Expiration date: 2024-05-27

Components	Quantity	Lot	Expiration Date
C1	6 vials	FA2259AV1D	2024-05-27
C2	6 vials	FA2259AW1D	2024-05-27

Analytical Data	Specification	Result
Concentration	C1: 50 -200 ng/mL C2: 200 -400 ng/mL	C1: 112 ng/mL C2: 337 ng/mL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2021-12-22

Quality Control Manager: S. LECOURT