

 HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN RIVAROXABAN CALIBRATOR LOW

Product Code : 226001

Lot No. : FA22971A

Expiration date : 2024-06-04

| Components | Quantity | Lot | Expiration Date |
|------------|----------|------------|-----------------|
| CAL I | 4 vials | FA2297DJ1A | 2024-06-04 |
| CAL II | 4 vials | FA2297DK1A | 2024-06-04 |
| CAL III | 4 vials | FA2297DL1A | 2024-06-04 |

| Analytical Data | Specification | Result |
|--|--|----------------------------------|
| Concentration | CAL I : < 20 ng/ml CAL II : 40 -60 ng/ml CAL III : 80 -120 ng/ml | 0 ng/mL 54 ng/mL 103 ng/mL |
| Linearity and accuracy of calibration with values assigned | $R^2 \geq 0.98$ Measured values for controls in acceptance ranges | PASS |

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2022-01-17

Quality Control Manager : S.LECOURT

