



HYPHEN BioMed
 155 rue d'Eragny
 95000 Neuville-sur-Oise
 France

CERTIFICATE OF ANALYSIS

Product Description : Factor IX Deficient Plasma
Product Code : DP050A/K
Lot No. : FB00141G
Expiration date : 2024-06-24

Components	Quantity	Lot	Expiration Date
Factor IX Deficient Plasma	1 vial (DP050A) 6 vials (DP050K)	FB0014JW1G	2024-06-24

Analytical Data	Specification	Result
Calibration curve	CT (0-100%) \geq 40 sec CT (100%) $<$ 80 sec	91 sec 48 sec
Concentration FIX	$< 1\%$	$< 1\%$
Concentration FV	$> 50 \%$	95%

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2022-02-02

Quality Control Manager : S.LECOURT

