



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description: BIOPHEN HEPARIN CALIBRATOR
Product Code: 222001
Lot No.: FB00081D
Expiration date: 2024-07-27

Components	Quantity	Lot	Expiration Date
CAL 1	4 vials	FB0008BJ1D	2024-07-27
CAL 2	4 vials	FB0008BK1D	2024-07-27
CAL 3	4 vials	FB0008BL1D	2024-07-27
CAL 4	4 vials	FB0008BM1D	2024-07-27
CAL 5	4 vials	FB0008BN1D	2024-07-27

Analytical Data	Specification	Result
Concentration	CAL 1 : ≤ 0.05 IU/mL CAL 2 : 0.30 -0.50 IU/mL CAL 3 : 0.70 -0.90 IU/mL CAL 4 : 1.10 -1.30 IU/mL CAL 5 : 1.45 -1.75 IU/mL	0 0.36 0.78 1.14 1.50
Linearity and accuracy of calibration with values assigned	$R^2 \geq 0.98$ Measured values for controls in acceptance ranges	PASS

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2022-03-29

Quality Control Manager:

S.LECOURT
