



A Company of Sysmex Group



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description:** BIOPHEN RIVAROXABAN CONTROL PLASMA  
**Product Code:** 224501  
**Lot No.:** FB01061B  
**Expiration date:** 2024-07-23

Components	Quantity	Lot	Expiration Date
C1	6 vials	FB0106BD1B	2024-07-23
C2	6 vials	FB0106BE1B	2024-07-23

Analytical Data	Specification	Result
Concentration	C1: 50 -150 ng/mL C2: 250 -350 ng/mL	C1: 98 ng/mL C2: 313 ng/mL

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval: 2022-02-17

Quality Control Manager:

S.LECOURT