

 HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description:** BIOPHEN APIXABAN CALIBRATOR  
**Product Code:** 226201  
**Lot No.:** FB01091A  
**Expiration date:** 2024-07-06

Components	Quantity	Lot	Expiration Date
CAL 1	4 vials	FB0109DO1A	2024-07-06
CAL 2	4 vials	FB0109DP1A	2024-07-06
CAL 3	4 vials	FB0109DQ1A	2024-07-06

Analytical Data	Specification	Result
Concentration	CAL 1: $\leq 50$ ng/ml CAL 2: 200 -400 ng/ml CAL 3: 500 -700 ng/ml	0 317 632
Linearity and accuracy of calibration with values assigned	$R^2 \geq 0.98$ Measured values for controls in acceptance ranges	PASS PASS

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval: 2022-02-15

Quality Control Manager: S. LECOURT

