



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description: CEPHEN 5
Product Code: CK515K/L
Lot No.: FB1181AN
Expiration date: 2024-07-21

Components	Quantity	Lot	Expiration Date
CEPHEN 5	8 vials (CK515K) 12 vials (CK515L)	FB1181KBAN	2024-07-21

Analytical Data	Specification	Result
aPTT on Normal Plasma	Ratio < 1.20	1.05 (CS - CN) 1.02 (STAR)

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2022-08-23

Quality Control Manager: S. LECOURT

