



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description: BIOPHEN EDOXABAN CALIBRATOR LOW
Product Code: 226401
Lot No.: FB037618
Expiration date: 2024-08-13

Components	Quantity	Lot	Expiration Date
CAL I	4 vials	FB0376af18	2024-08-13
CAL II	4 vials	FB0376ag18	2024-08-13
CAL III	4 vials	FB0376ah18	2024-08-13

Analytical Data	Specification	Result
Concentration	CAL I : < 15 ng/mL CAL II : 40 -60 ng/mL CAL III : 80 -120 ng/mL	0 55 113
Linearity and accuracy of calibration with values assigned	$R^2 \geq 0.98$ Measured values for controls in acceptance ranges	PASS PASS

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2022-03-29

Quality Control Manager: S.LECOURT