



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description: COLLAGEN
Product Code: AG005K
Lot No.: FB0330-FB0331
Expiration date: 2024-08-17

Components	Quantity	Lot	Expiration Date
COLLAGEN REAGENT	3 vials	FB0330-FB0331	2024-08-17
COLLAGEN DILUENT	3 vials	FB0330-FB0331	2024-08-23

Analytical Data	Specification	Result
Reactivity (% AgregMax) at 2 µg/ml on CS	≥ 70%	100%
Reactivity (% AgregMax) at 2 µg/ml on Agregometer	≥ 70%	96%

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2022-03-29

Quality Control Manager: S. LECOURT