



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description: ZYMUTEST FACTOR VII
Product Code: RK036A
Lot No.: FB0352
Expiration date: 2024-08-30

Components	Quantity	Lot	Expiration Date
Anti (h) Factor VII pre-coated plate	12x8 wells	FB0479	2024-09-01
Factor VII standard	3 vials	FB0478	2024-08-30
Factor VII Control I	1 vial	FB0477	2025-08-25
Factor VII Control II	1 vial	FB0476	2025-08-25
Anti-(h) Factor VII HRP immunoconjugate	3 vials	FB0512	2024-09-06
Sample diluent	2x50 ml	FA1738	2025-03-10
Wash solution	1x50 ml	FA2211	2025-05-12
Conjugate diluent	1x25 ml	FA1521	2025-02-24
TMB	1x25 ml	210302D01	2025-03-30
Sulfuric Acid 0,45M	1x6 ml	FA1514	2025-02-17

Analytical Data	Specification	Result
Reactivity	A450 for 100 % ≥ 1.00	1.28
Calibration Curve and Accuracy	Controls Measured Values within acceptance ranges	Controls Measured Values within acceptance ranges
Blank	< 0.100	0.015
Detection threshold	$\leq 5\%$	$< 1\%$

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2022-04-13

Quality Control Manager: S. LECOURT