



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description: BIOPHEN HEPARIN (AT+)
Product Code: 221007
Lot No.: FB054118
Expiration date: 2024-09-14

Components	Quantity	Lot	Expiration Date
R1 : Sxa-11 substrate	2 vials	FB0541HO18	2024-09-14
R2 : Bovine Fxa	2 vials	FB0541HP18	2024-10-05
R3 : Human AT	2 vials	FB0541HQ18	2024-09-20

Analytical Data	Specification	Result
SUPERIMPOSITION REACTIVITY, LINEARITY, ACCURACY ON STAR	<ul style="list-style-type: none">• For a same A405 (UFH/LMWH): MV = TV \pm0.05 IU/ml for conc. \leq 0.50 IU/ml MV = TV \pm0.10 IU/ml for conc. $>$ 0.50 IU/ml• R² \geq 0.98	PASS
DETECTION THRESHOLD	\leq 0.05 IU/mL	$<$ 0.01 IU/mL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2022-05-31

Quality Control Manager: S.LECOURT