



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

<b>Product Description:</b>	<b>BIOPHEN EDOXABAN CALIBRATOR</b>
<b>Product Code:</b>	<b>226501</b>
<b>Lot No.:</b>	<b>FB06081A</b>
<b>Expiration date:</b>	<b>2024-09-24</b>

Components	Quantity	Lot	Expiration Date
CAL 1	4 vials	FB0608aa1A	2024-09-24
CAL 2	4 vials	FB0608ab1A	2024-09-24
CAL 3	4 vials	FB0608ac1A	2024-09-24

Analytical Data	Specification	Result
Concentration	CAL 1 : $\leq 40$ ng/mL CAL 2 : 200 -300 ng/mL CAL 3 : 400 -600 ng/mL	0 ng/mL 255 ng/mL 500 ng/mL
Linearity and accuracy of calibration with values assigned	$R^2 \geq 0.98$ Measured values for controls in acceptance ranges	PASS PASS

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval: 2022-05-02

Quality Control Manager: S.LECOURT

