



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description: HEMOCLOT QUANTI. V-L
Product Code: CK065K
Lot No.: FB06901B
Expiration date: 2024-10-07

Components	Quantity	Lot	Expiration Date
R1: coagulation factors	3 vials	FB0690FS1B	2024-10-07
R2 : Purified Factor Xa-PLPs	3 vials	FB0690FT1B	2024-10-25

Analytical Data	Specification	Result
Calibration curve (manual method)	Delta CT (100-10%) ≥ 35 sec	56 sec
	CT (100%) = 32 ± 5 sec	32 sec
	$R^2 \geq 0.98$	PASS
	Controls measured in acceptance ranges	PASS
	Normal Plasma $\leq 10\%$	PASS

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2022-06-08

Quality Control Manager: S. LECOURT