



A Company of Sysmex Group



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description:** BIOPHEN ARIXTRA CONTROL PLASMA  
**Product Code:** 224001  
**Lot No.:** FB08551A/FB085414  
**Expiration date:** 2024-11-02

Components	Quantity	Lot	Expiration Date
C1	6 vials	FB0855CR1A/FB085414	2024-11-02
C2	6 vials	FB0855CS1A/FB085414	2024-11-02

Analytical Data	Specification	Result
Concentration	C1 : 0.30 -0.50 µg/mL C2: 1.05 -1.35 µg/mL	0.41 1.22

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval: 2022-06-22

Quality Control Manager:

S.LECOURT