



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description: BIOPHEN APIXABAN CALIBRATOR
Product Code: 226201
Lot No.: FB09011B
Expiration date: 2024-11-19

Components	Quantity	Lot	Expiration Date
CAL 1	4 vials	FB0901DO1B	2024-11-19
CAL 2	4 vials	FB0901DP1B	2024-11-19
CAL 3	4 vials	FB0901DQ1B	2024-11-19

Analytical Data	Specification	Result
Concentration	CAL 1: ≤ 50 ng/ml CAL 2: 200 -400 ng/ml CAL 3: 500 -700 ng/ml	0 304 526
Linearity and accuracy of calibration with values assigned	$R^2 \geq 0.98$ Measured values for controls in acceptance ranges	PASS PASS

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2022-06-28

Quality Control Manager: S. LECOURT

