



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description: BIOPHEN ORGARAN CONTROL

Product Code: 223501

Lot No.: FB09841B

Expiration date: 2024-12-03

Components	Quantity	Lot	Expiration Date
C1	6 vials	FB0984CL1B	2024-12-03
C2	6 vials	FB0984CM1B	2024-12-03

Analytical Data	Specification	Result
Concentration	C1 : 0.40 -0.60 U/mL C2 : 0.80 -1.20 U/mL	0.58 U/mL 1.12 U/mL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2022-07-06

Quality Control Manager: S.LECOURT