



A Company of Sysmex Group



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description :** CEPHEN 1 LS

**Product Code :** CK521K

**Lot No. :** FB2046CH

**Expiration date :** 2024-12-22

Components	Quantity	Lot	Expiration Date
CEPHEN 1 LS	6 vials	FB2046KDCH	2024-12-22

Analytical Data	Specification	Result
aPTT on Normal Plasma	Ratio < 1.20	1.03

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval : 2023-01-19

Quality Control Manager :

**S. LECOURT**