



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description: BIOPHEN LMWH C4
Product Code: 224201
Lot No.: FB12622D
Expiration date: 2025-01-20

Components	Quantity	Lot	Expiration Date
C4	12 vials	FB1262BP2D	2025-01-20

Analytical Data	Specification	Result
Concentration	C4 : 1.05 -1.35 IU/mL	1.24 IU/mL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2022-08-31

Quality Control Manager: S. LECOURT