



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS	
Product Description :	ZYMUTEST Fibrinogen
Product Code :	RK024A
Lot No. :	FB1158
Expiration date :	2025-01-03

Components	Quantity	Lot	Expiration Date
pre-coated plate	12x8 wells	FB1231	2025-01-10
Fibrinogen Standard	3 vials	FB1173	2025-01-03
Fibrinogen Control I	1 vial	FB1174	2025-01-03
Fibrinogen Control II	1 vial	FB1175	2025-01-03
Fibrinogen HRP immunoconjugate	3 vials	FB1252	2025-01-20
B2F Sample diluent	2x50 ml	FB1253	2025-01-17
Wash solution	1x50 ml	FB0867	2025-11-03
B2F Conjugate diluent	1x25 ml	FB1254	2025-01-17
TMB	1x25 ml	210302D01	2025-03-30
Sulfuric Acid 0,45M	1x6 ml	FB0508	2025-09-01

Analytical Data	Specification	Result
Reactivity	A450 for 40 ng/ml \geq 1.45	2.062
Calibration Curve and Accuracy	Controls Measured Values within acceptance ranges	PASS
Blank	< 0.150	0.031
Detection threshold	\leq 1 ng/ml	< 0.5 ng/mL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2022-08-25

Quality Control Manager: S.LECOURT

