



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN APIXABAN CONTROL LOW
Product Code : 225201
Lot No. : FB15371F
Expiration date : 2025-03-15

Components	Quantity	Lot	Expiration Date
C1	6 vials	FB1537DW1F	2025-03-15
C2	6 vials	FB1537DX1F	2025-03-15

Analytical Data	Specification	Result
Concentration	C1 : 15 -35 ng/ml C2 : 45 -105 ng/ml	26 ng/mL 80 ng/mL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2022-10-17

Quality Control Manager : S.LECOURT