



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : COLLAGEN
Product Code : AG005K
Lot No. : FB1631 / FB1630
Expiration date : 2025-03-31

Components	Quantity	Lot	Expiration Date
COLLAGEN REAGENT	3 vials	FB1631 / FB1630	2025-03-31
COLLAGEN DILUENT	3 vials	FB1631 / FB1630	2025-04-04

Analytical Data	Specification	Result
Reactivity (% AgregMax) at 2 µg/ml on CS	≥ 70%	96%
Reactivity (% AgregMax) at 2 µg/ml on Agregometer	≥ 70%	101%

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2022-10-27

Quality Control Manager : **S.LECOURT**