



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : HEMOCLOT QUANTI. V-L
Product Code : CK065K
Lot No. : FB15851C
Expiration date : 2025-03-23

Components	Quantity	Lot	Expiration Date
R1: coagulation factors	3 vials	FB1585FS1C	2025-03-23
R2 : Purified Factor Xa-PLPs	3 vials	FB1585FT1C	2025-03-30

Analytical Data	Specification	Result
Calibration curve (manual method)	Delta CT (100-10%) ≥ 35 sec	77 sec
	CT (100%) = 32 ± 5 sec	32 sec
	$R^2 \geq 0.98$	
	Controls measured in acceptance ranges	PASS
	Normal Plasma $\leq 10\%$	PASS

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2022-10-25

Quality Control Manager :

S.LECOURT