



 HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description :** COLLAGEN  
**Product Code :** AG005K  
**Lot No. :** FC2144  
**Expiration date :** 2026-03-27

Components	Quantity	Lot	Expiration Date
COLLAGEN REAGENT	3 vials	FC2144	2026-03-31
COLLAGEN DILUENT	3 vials	FC2144	2026-03-27

Analytical Data	Specification	Result
Reactivity (% AgregMax) at 2 µg/ml on CS or CN Series	≥ 70%	86%
Reactivity (% AgregMax) at 2 µg/ml on Agregometer	≥ 70%	79%

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval : 2023-11-02

Quality Control Manager : S. LECOURT