



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description :** BIOPHEN DABIGATRAN PLASMA CALIBRATOR LOW  
**Product Code :** 222901  
**Lot No. :** FC239619  
**Expiration date :** 2026-05-16

Components	Quantity	Lot	Expiration Date
CAL 1	4 vials	FC2396DA19	2026-05-16
CAL 2	4 vials	FC2396DB19	2026-05-16
CAL 3	4 vials	FC2396DC19	2026-05-16

Analytical Data	Specification	Result
Concentration	CAL 1 : < 20 ng/ml CAL 2 : 40 -60 ng/ml CAL 3 : 80 -110 ng/ml	0 ng/mL 55 ng/mL 107 ng/mL
Linearity and accuracy of calibration with values assigned	$R^2 \geq 0.98$ Measured values for controls in acceptance ranges	PASS PASS

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

### Quality Control Approval

Date of Approval : 27 FEB. 2024

Quality Control Manager :

S. LECOURT