



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS	
Product Description :	ARGATROBAN CONTROL PLASMA
Product Code :	SC035K
Lot No. :	FD00621A
Expiration date :	2026-06-30

Components	Quantity	Lot	Expiration Date
C1	6 vials	FD0062CE1A	2026-06-30
C2	6 vials	FD0062CF1A	2026-06-30

Analytical Data	Specification	Result
C1	0.55 - 0.80 µg/mL	0.75 µg/mL
C2	1.10 - 1.40 µg/mL	1.38 µg/mL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval :

Quality Control Approval

Quality Control Manager :

06 FEB. 2024
S. LECOURT

