




A Company of Sysmex Group

 HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN ANTITHROMBIN 5
Product Code : 221105
Lot No. : FD0256HX
Expiration date : 2026-07-20

Components	Quantity	Lot	Expiration Date
R1 : Bovin Xa	4 vials	FD0256FQHX	2026-07-27
R2 : Substrate	4 vials	FD0256FRHX	2026-07-20
R3: Tris heparin Buffer	4 vials	FD0256cyHX	2026-07-26

Analytical Data	Specification	Result
Calibration curve : reactivity, linearity, slope and accuracy	A405 (0%) = 1.50 ± 0.20 Delta A405 (0-100%) ≥ 0.80 $R^2 \geq 0.98$ Measured values for controls in acceptance ranges	1.52 1.04 0.99 PASS
Detection Threshold	$\leq 5\%$	< 4 %

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Quality Control Approval

Date of Approval :

04 MARS 2024

S. LECOURT

Quality Control Manager :