



Manufactured By: HYPHEN BioMed

Fibrinogen Fragment E

Fragment E du Fibrinogène

Vial of 200 µg/flacon de 200 µg

Ref. APP018A

STORE AT
2-8°C

For research use only



Uniquement à usage de
recherche

CONSERVER A
2-8°C

Origin:

Highly purified fibrinogen Fragment E, obtained by degradation of Fibrinogen with plasmin, then purification with ion exchange and gel filtration chromatographies.

Presentation:

Vial containing approximately 200 µg of highly purified human fibrinogen, fragment E.

Reconstitution:

Each vial must be restored with 1 ml distilled water. It can be diluted to the desired concentration when required. Use a buffer containing a carrier protein (such as BSA) for dilution, in order to avoid material loss.

Excipients:

- Glycine, trisodium citrate and Hepes.
- 9 mg sodium chloride/vial

Purity:

One major band of about 50,000 daltons on SDS-PAGE.

Viral safety:

Fibrinogen used for preparing fragment E was extracted from human plasma, tested with registered methods and found negative for HIV antibodies, HBs Ag and HVC antibodies. However, no assay may warrant the total absence of infectious agents. Any product of human origin must then be handled with all the required cautions, as being potentially infectious.

Use:

For in vitro use only.
All research studies and protocols where a source of highly purified fibrinogen is necessary.

Stability:

- Restored :
- 72 hours at room temperature
 - 7 days at 2-8 °C
 - 6 months frozen at -20 °C or below

Lyophilized : Until the expiration date printed on the vial.

Origine :

Protéine hautement purifiée obtenue par clivage de fibrinogène humain purifié par la plasmine et purification par chromatographie d'échange d'ions et de filtration sur gel.

Présentation :

Flacon contenant environ 200 µg de fragment E du fibrinogène humain.

Reconstitution:

Chaque flacon doit être reconstitué par 1 ml d'eau distillée. Il peut être dilué à la concentration requise, si nécessaire. Utiliser un tampon contenant un adjuvant protéique (comme la BSA) afin d'éviter toute perte protéique par adsorption sur les parois du récipient.

Excipients :

- Glycine, trisodium citrate et hépes.
- 9 mg de chlorure de sodium/flacon

Pureté :

Bande principale d'environ 50 000 daltons sur SDS-PAGE.

Sécurité virale :

Le fibrinogène servant à la préparation du fragment E a été extrait de plasma humain et testé par des méthodes enregistrées et est certifié exempt d'anticorps VIH, de Hbs Ag et d'anticorps VCH. Toutefois, aucune méthode ne permettant d'exclure totalement le risque d'agent pathogène, ces produits doivent être manipulés avec toutes les précautions requises pour l'utilisation de produits potentiellement infectés.

Utilisation :

Utilisation *in vitro* exclusivement.
Toutes les études ou protocoles dans lesquels l'utilisation d'une source de fibrinogène hautement purifié est nécessaire.

Stabilité du flacon :

- Reconstitué :
- 72 heures à température ambiante
 - 7 jours à 2-8 °C
 - 6 mois congelé à -20 °C ou moins

Lyophilisé : Jusqu'à la date de péremption indiquée sur le flacon.

Fibrinogen Fragment E

Fragment E du Fibrinogène

Vial of 200 µg/flacon de 200 µg

Ref. APP018A

STORE AT
2-8°C

Lot :



Expiration :

CONSERVER A
2-8°C

ANALYSIS CERTIFICATE

	ANALYTICAL DATA	SPECIFICATIONS
1.	Protein Content (Lowry method; A280 nm) <div style="text-align: center;">..... µg/vial</div>	> 180 µg/vial
2.	SDS-PAGE (5 % acrylamide) 1 single band of about daltons	1 major band of about 50,000 daltons

CERTIFICAT D'ANALYSE

	RESULTATS D'ANALYSE	SPECIFICATIONS
1.	Teneur protéique (Méthode Lowry; D0280 nm) <div style="text-align: center;">..... µg/flacon</div>	> 180 µg/flacon
2.	SDS-PAGE (5 % acrylamide) 1 bande d'environ daltons	1 bande principale ≈ 50 000 daltons

DATE :

CONCLUSIONS : Passed/Conforme Refused/Non-conforme

Quality Control Manager
Responsable contrôle qualité