

ZYMUTEST Anti- β_2 -GLYCOPROTEIN I IgG - Isotype # ARK014A-RUO

Auto-antibodies to β_2 GlycoProtein I (β_2 GPI), IgG isotype

**FOR RESEARCH USE ONLY.
NOT FOR USE IN DIAGNOSTIC PROCEDURES.**



Last revision : 30/03/2011

INTENDED USE:

The ZYMUTEST anti- β_2 GPI, IgG ELISA kit, is a standardised and optimised enzyme immuno-assay designed for measuring auto-antibodies to β_2 GPI of the IgG isotype, in human plasma or serum or in any biological fluid where auto-antibodies to β_2 GPI must be measured. **This assay is for research use only and should not be used for patient diagnosis or treatment.**

ASSAY PRINCIPLE:

The diluted plasma sample or biological fluid is introduced into one of the microwells of the β_2 GPI coated plate. When present, anti- β_2 GPI auto-antibodies bind to immobilised β_2 GPI. Following a washing step, bound auto-antibodies, of the IgG isotype, are revealed with a goat anti-human IgG (Fc γ specific)-peroxidase conjugate, which reacts specifically with IgG isotypes. Following a new washing step, the peroxidase substrate, Tetramethylbenzidine (TMB) in presence of hydrogen peroxide (H₂O₂), is introduced and a blue colour develops. The colour turns yellow when the reaction is stopped with sulfuric acid. The colour developed is directly proportional to the amount of anti- β_2 GPI auto-antibodies, of the IgG isotype, present in the tested sample.

TESTED SAMPLES:

- Trisodium citrate or Na₂ EDTA anticoagulated human plasma or human serum.
- Any biological fluid, where human auto-antibodies to β_2 GPI, of the IgG isotype, must be assayed.

REAGENTS:

1. **COAT:** Micro ELISA plate, containing 12 strips of 8 wells, coated with highly purified human β_2 GPI, then stabilized; the plate is packed in an aluminium pouch hermetically sealed in presence of a desiccant.
2. **SD:** 2 vials containing 50 ml of **Autoimmunity Sample Diluent**, ready to use. Contains Sodium Azide.
3. **CAL:** 3 vials of **anti- β_2 GPI, IgG, calibrator**, lyophilised. When restored with 1 ml of **Autoimmunity Sample Diluent**, the ready to use calibrator is obtained (already diluted 1:100).

Note: This calibrator has a defined anti- β_2 GPI concentration, expressed in **Arbitrary Units (AU)** and indicated on the flyer provided with the kit.

4. **CC:** 3 vials of **negative control**, lyophilised (diluted normal human plasma). When restored with 1 ml of **Autoimmunity Sample Diluent**, the ready to use negative control is obtained (already diluted 1:100).
5. **IC:** 3 vials of **immunoconjugate (Anti-IgG (Fc γ)-HRP immunoconjugate)**, affinity purified goat antibodies specific for human IgG-Fc γ coupled to HRP, lyophilised.
6. **CD:** 1 vial of 25 ml of **Conjugate Diluent**, ready to use.
7. **WS:** 1 vial of 50 ml of 20 fold concentrated **Wash Solution**.
8. **TMB:** 1 vial of 25 ml peroxidase substrate: 3,3',5,5' - **Tetramethylbenzidine (TMB)** containing hydrogen peroxide, ready to use.
9. **SA:** One vial of 6 ml of 0.45M **Sulfuric acid (Stop Solution)**, ready to use.

Note: Use only components from a same kit lot. Do not mix components from different lots of kits, when running the assay.

REAGENTS AND EQUIPMENT REQUIRED BUT NOT PROVIDED:

- 8-channel or repeating pipette allowing dispensing 50-300 μ l.
- 1-channel pipettes at variable volumes from 0 to 20 μ l, 20 to 200 μ l and 200 to 1000 μ l.
- Micro ELISA plate washing equipment and shaker.
- Micro ELISA plate reader with a wavelength set up at 450 nm.
- Distilled water.

REAGENTS PREPARATION, STORAGE AND STABILITY:

In their original packaging box, before use, when stored at 2-8°C, the unopened reagents are stable until the expiration date printed on the box.

1. **Micro ELISA plate:** open the plastic pouch and take off the required amounts of 8 well strips for the test series. When out of the pouch, the strips must be used within 30 minutes. Unused strips can be stored at **2-8°C for 4 weeks** in their original aluminium pouch, in presence of the desiccant, hermetically closed and protected from any moisture, and stored in the provided microplate storage bag (minigrip).
2. **Autoimmunity Sample Diluent:** It is ready to use. When open, it can be used for **4 weeks**, stored at **2-8°C**, and provided that any bacterial contamination is avoided during use. This reagent contains sodium azide.

Warning: The Autoimmunity Sample Diluent contains sodium azide, which may react with lead and copper plumbing to form highly explosive metal azides. Flush with large volumes of water when discarding into a sink.

3. **Calibrator:** restore each vial with 1 ml Autoimmunity Sample Diluent in order to obtain the ready to use calibrator. It corresponds to a plasma containing IgG isotype auto-antibodies to β_2 GPI, already **diluted 1:100**. Following reconstitution, the calibrator is stable for **5 days at 2-8°C**, provided any bacterial contamination is avoided during use.
4. **Negative control:** restore each vial with 1 ml Autoimmunity Sample Diluent in order to obtain the ready to use negative control. It corresponds to a normal human plasma, already **diluted 1:100**. Following reconstitution, the negative control is stable for **2 weeks at 2-8°C**, provided any bacterial contamination is avoided during use.

Warning: β_2 GPI used for coating the plates is extracted from human plasma. Negative control is also prepared with human plasma. Any human plasma used is tested with registered methods and found negative for HIV antibodies, HBS Ag and HCV antibodies. However, no assay may warrant the total absence of infectious agents. Any product of human origin must then be handled with all the required cautions, as being potentially infectious.

5. **Anti-IgG (Fc γ)-HRP immunoconjugate:** each vial must be restored with 7.5 ml of **Conjugate Diluent**. Let the pellet to be completely dissolved before use, and shake the vial gently in order to homogenize the content. The restored conjugate is stable for at least **24 hours at room temperature** or for at least **4 weeks at 2-8°C**.
6. **Conjugate Diluent:** It is ready to use. When open, it can be used for **4 weeks**, stored at **2-8°C**, and provided that any bacterial contamination is avoided during use. This reagent contains 0.05% Kathon CG.
7. **Wash Solution:** Incubate the vial for 15-30 minutes in a water bath, at 37°C, until complete dissolution of solids, when present. Shake the vial and dilute the amount required 1:20 in distilled water (the 50 ml contained in the vial allow to prepare 1 liter of Wash Solution). The Wash Solution must be stored at **2-8°C** in its original vial and used within **4 weeks following opening**. The diluted Wash Solution must be used within **7 days**, when protected from any contamination. This reagent contains 0.05% Kathon CG.
8. **TMB substrate:** It is ready to use. When open, it can be used for **4 weeks**, stored at **2-8°C**, and provided that any bacterial contamination is avoided during use.
9. **Stop solution:** It is ready to use.

Cautions: Sulfuric acid, although diluted to 0.45M is caustic. As for any similar chemical, handle Sulfuric acid with great care. Avoid any skin and eye contact. Wear protection glasses and gloves when handling.

Note: Bring the kit at room temperature, at least 30 min. before use. Store the unused reagents at 2-8°C. The stability studies at 30°C show that the reagents can be shipped at room temperature without damage.

PROCEDURE:

Sample collection:

Blood plasma (9 vol.) must be collected on 0.109 M citrate anticoagulant (1 vol.); plasma supernatant is decanted following a 20 min. centrifugation at 2,500 g; citrated plasma should be tested within **24 hours** or stored frozen at -20°C or below for up to 6 months, and thawed for 15 min. at 37°C just before use. Thawed specimen must be tested within **12 hours**. EDTA collected human plasma may also be used.

Auto-antibodies to β_2 GPI can also be assayed on serum.

Tested plasma or sample or control:

Plasma or serum is tested at **1:100** dilution in Autoimmunity Sample Diluent. When high amounts of auto-antibodies to β_2 GPI are expected, samples must be assayed at **1:200** or **1:400** dilutions. Results must then be multiplied by 2 or 4.

Calibrator and negative control are ready to use (**already diluted 1:100**).



7768 Service Center Drive • West Chester OH 45069

Phone: 513.770.1991

Toll Free: 866.783.3797

Fax: 513.573.9241

Email: info@aniara.com

D.750.02/ZY/014A/RUO

www.aniara.com

Assay procedure:

Calibration curve: The assay can be calibrated with the calibrator provided in the kit, and which concentration (C) is indicated in arbitrary units, (AU) on the flyer provided. Prepare the standard solutions for calibration by doing a serial two-step dilution of the calibrator in Autoimmunity Sample Diluent, from 1:1 to 1:32. A concentration range from C:1 to C:32 is obtained.

Remove the required number of strips from the aluminium pouch, for the series of measures to be performed. Then put the strips in the frame provided. In the different wells of the micro ELISA plate introduce the reagents and perform the various assay steps as indicated on the following table:

Reagent	Volume	Procedure
Anti- β_2 GPI IgG Calibrator or Negative control or 1:100 diluted sample or sample diluent (blank)	200 μ l	Introduce the : – Calibrator or – negative control or – diluted sample or – sample diluent into the micro ELISA plate wells.
Incubate for 30 minutes at room temperature (18-25 °C) (a) (b)		
Wash Solution (20 fold diluted in distilled water)	300 μ l	Proceed to 5 successive washings using the washing instrument. (b)
Conjugate anti-IgG (Fc γ)-HRP immunoconjugate, restored with 7.5 ml of conjugate diluent	200 μ l	Immediately after the washing, Introduce the anti-IgG (Fc γ)-HRP immunoconjugate in the micro ELISA plate wells.
Incubate for 30 minutes at room temperature (18-25 °C) (a)		
Wash Solution (20 fold diluted in distilled water)	300 μ l	Proceed to 5 successive washings using the washing instrument (b)
TMB/H ₂ O ₂ Substrate	200 μ l	Immediately after the washing, introduce the substrate into the wells. Note: The substrate distribution, row by row, must be accurate and at exact time intervals (c).
Let the colour develop for 5 min. at room temperature (18-25 °C) (a)		
0.45M Sulfuric Acid	50 μ l	Following exactly the same time intervals than for the addition of substrate, stop the colour development by introducing the 0.45M sulfuric acid (c)
Wait for 10 minutes in order to allow the colour to stabilize and measure absorbance at 450 nm (A ₄₅₀) (d). Subtract the blank value		

Note:

- Avoid letting the plate in the bright sunlight during incubations and more particularly during colour development. A micro ELISA plate shaker can be used.
- Never let the plates empty between the addition of the reagents or following the washing step. The next reagent must be added within 3 minutes, in order to prevent the plate from drying, which could damage the immobilized components. If necessary, keep the plate filled with Wash Solution and empty it just before the introduction of the next reagent. The washing instrument must be adjusted in order to wash the plates gently, and to avoid a too drastic emptying, which could lower plate reactivity.
- For addition of the TMB substrate, the time interval between each row must be accurate and exactly determined. It must be the same when stopping the reaction.
- For bichromatic readings, a reference wavelength at 690 nm or at 620 nm can be used.

QUALITY CONTROL:

- Calibrator and controls provided in the kit allow validating the right performance of the assay.
- Expected A₄₅₀ values for undiluted calibrator and the negative control can present variations from lot to lot but they always are:

$$P = A_{450} \text{ for 1:1 calibrator} : \geq 1.5$$

$$N = A_{450} \text{ for negative control} : \leq 0.25$$

In addition, concentrations obtained for negative control must be within the acceptance range indicated on the flyer provided in the kit. If negative control is out of this range check carefully the assay conditions and re-run the assay, if required.

EXPRESSION OF RESULTS:

- Results are expressed according to the A₄₅₀ values obtained for samples, controls and using the calibration curve.
- The calibration curve is obtained by plotting the anti- β_2 GPI concentrations expressed in AU on the abscissae and the corresponding A₄₅₀ on the ordinates (see model on the flyer). The anti- β_2 GPI, autoantibody concentration, of the IgG isotype, for the sample, tested at the standard 1:100 dilution, and expressed in AU, is directly deduced from the curve.
- When higher dilutions are used, (i.e. D), the concentration measured must be multiplied by the complementary dilution factor (i.e. D:100 ; for example x2 for 1:200 or x4 for 1:400).
- Alternatively, an ELISA software (i.e. Dynex, Biolise, etc...), can be used for the calculation of concentrations.
- The results obtained should be for research purposes only and not used for patient diagnosis or treatment.

LIMITATIONS OF THE ASSAY:

If the washing step is not correctly performed, the negative control can produce a high absorbance value. In order to avoid non-specific colour development, check that the washing step is performed efficiently.

As for any auto-antibody assay, clinical situation such as presence of inflammation, infectious diseases, auto-immune diseases, immun-complexes, high concentrations of IgG in the tested sample, can induce a high background, which can be within the grey zone or in the weak positive range. Check then for the possible presence of antibodies on another specimen collected later.

ASSAY SPECIFICITY AND CHARACTERISTICS:

The ZYMUTEST anti- β_2 GPI, IgG Kit, specifically measures human autoantibodies to β_2 GPI of the IgG isotype, reactive with immobilised β_2 GPI. IgM or IgA isotypes are not measured.

This assay is designed with native uncleaved and non-altered, highly purified human β_2 GPI, which has then a preserved structure. This method then provides high reproducibility, high sensitivity and high specificity.

REFERENCES:

- Arvieux J., Roussel B., Jacob M.C., Colomb M.G. : Measurement of antiphospholipid antibodies by ELISA using β_2 -Glycoprotein I as antigen. J. Immunol. Meth, 1991, 143, 223-229.
- Viard J.P., Amoura Z., Bach J.F. : Association of Anti- β_2 -Glycoprotein I Antibodies with Lupus Type Circulating Anticoagulant and Thrombosis in Systemic Lupus Erythematosus. Am. J. Med., 1992, 93, 181-86.
- Martinuzzo M.E., Forastiero R.P., Carreras L.O. : Anti- β_2 -Glycoprotein I antibodies : detection and association with Thrombosis. Brit. J. haemat., 1995, 89, 397-402.
- Sanmarco M., Soler C. : Heterogeneity of β_2 -Glycoprotein I antibodies. Nouv. Rev. Fr. Haemat., 1995, 37, S57-S60.
- Amengual O., Atsumi T., Khamashta M.A., Hughes G.R.V. : Clinical significance of anti- β_2 -Glycoprotein I antibodies. Am. Med. Interne, 1997, 147-1, 15-17.