

# Purified Human Fibrinogen

## Fibrinogène humain purifié

Vial of 100 mg/flacon de 100 mg

STORE AT  
2-8 °C

CONSERVER A  
2-8 °C

Ref. PP001B

Lot: 140407A

Expiration: 2018-04



### ANALYSIS CERTIFICATE

### CERTIFICAT D'ANALYSE

ANALYTICAL DATA	SPECIFICATIONS
1. <u>Protein Content</u> (Lowry method ) 113.6 mg/vial	> 90 mg/vial
2. <u>SDS-PAGE</u> (5 % acrylamide) 1 single band of about 340 000 daltons	1 major band of about 340,000 daltons
3. <u>Clotting time</u> (5 NIH/ml Thrombin) 5.7 sec	< 15 sec
4. <u>Coagulability</u> 98.8 %	≥ 98%

RESULTATS D'ANALYSE	SPECIFICATIONS
1. <u>Teneur protéique</u> (Méthode Lowry ) 113.6 mg/flacon	> 90 mg/flacon
2. <u>SDS-PAGE</u> (5 % acrylamide) 1 bande d'environ 340 000 daltons	1 bande principale ≈ 340 000 daltons
3. <u>Temps de coagulation</u> (Thrombine 5 NIH/ml) 5.7 sec	< 15 sec
4. <u>Coagulabilité</u> 98.8 %	≥ 98%

DATE : 05/05/2014      CONCLUSIONS :  Passed/Conforme     Refused/Non-conforme

Quality Control Manager  
Responsable contrôle qualité  
S. LECOURT

po ssc