

ANALYSIS CERTIFICATE

BIOPHEN FVIII:C- #221402

Lot : F1500002

QC Release : 20/01/2016

Expiration date : 2018-05-19

Components	Volume	Exp. (months)	Int. Ref.	Lot #	Exp. date
R1 : Human Factor X	2 vials	30	NA	F151100002	2018-05-21
R2 : Activation Reagent	2 vials	30	NA	F151100002	2018-05-27
R3 : SXa-11 substrate	2 vials	30	NA	F151100002	2018-05-21
R4+ : Tris-BSA buffer	4 vials	30	NA	F151100002	2018-05-19

305

ANALYSIS CERTIFICATE

BIOPHEN FVIII:C- #221402

Lot : F150002

QC Release : 20/01/2016

Expiration date : 2018-05-19

Analytical data	Specifications				
<p>1. <u>Human Factor X</u></p> <p>a. Reproducibility (water bath)(100 % VIII:C) N = 20 Mean (A405): 1,278 CV: 1,47 %</p> <p>b. Factor X concentration (water bath) 81 %</p>	<p>N ≥ 10 ≤ 2 % ≥ 50 %</p>				
<p>2. <u>Activation Reagent (IXa - Thrombin - PLPs - calcium)</u></p> <p>a. Reproducibility (water bath)(100 % VIII:C) N = 20 Mean (A405): 1,257 CV: 1,36 %</p>	<p>N ≥ 10 ≤ 2 %</p>				
<p>3. <u>SXa-11 substrate</u> (tested at 3 mg/ml for a,b)</p> <p>a. Blank value (N=10) Mean (A405): 0,229</p> <p>b. Stability of substrate blank (A405)</p> <table border="1" style="margin-left: auto; margin-right: auto;"> <thead> <tr> <th style="text-align: center;">Fresh</th> <th style="text-align: center;">7 days R.T</th> </tr> </thead> <tbody> <tr> <td style="text-align: center;">0,221</td> <td style="text-align: center;">0,232</td> </tr> </tbody> </table> <p>c. Reproducibility (water bath) N = 20 Mean (A405): 1,180 CV: 1,54 %</p>	Fresh	7 days R.T	0,221	0,232	<p>A405 ≤ 0.30 A405 ≤ 0.30 N ≥ 10 ≤ 2 %</p>
Fresh	7 days R.T				
0,221	0,232				

ANALYSIS CERTIFICATE

BIOPHEN FVIII:C- #221402

Lot : F1500002

QC Release : 20/01/2016

Expiration date : 2018-05-19

Analytical data		Specifications																																																					
1. Calibration curve High Range: <table border="1"> <thead> <tr> <th colspan="2">VIII:C (%)</th> <th>A405 W. Bath</th> <th>A405 STAR</th> </tr> </thead> <tbody> <tr> <td>0</td> <td>0/0</td> <td>0,015</td> <td>0</td> </tr> <tr> <td>C/4</td> <td>25/20</td> <td>0,246</td> <td>0,209</td> </tr> <tr> <td>C/2</td> <td>50/50</td> <td>0,584</td> <td>0,442</td> </tr> <tr> <td>C</td> <td>100/100</td> <td>1,272</td> <td>0,682</td> </tr> <tr> <td>2C</td> <td>200/190</td> <td>2,323</td> <td>0,759</td> </tr> <tr> <td colspan="2">Linearity: r2 =</td> <td>0,9961</td> <td>0,9950</td> </tr> </tbody> </table> Low Range: <table border="1"> <thead> <tr> <th colspan="2">VIII:C (%)</th> <th>A405 W. Bath</th> <th>A405 STAR</th> </tr> </thead> <tbody> <tr> <td>0</td> <td>0/0.93</td> <td>0,011</td> <td>0,041</td> </tr> <tr> <td>C/4</td> <td>6,25/6,25</td> <td>0,294</td> <td>0,324</td> </tr> <tr> <td>C/2</td> <td>12,5/12,5</td> <td>0,628</td> <td>0,664</td> </tr> <tr> <td>C</td> <td>25/25</td> <td>1,234</td> <td>1,195</td> </tr> <tr> <td colspan="2">Linearity: r2 =</td> <td>0,9990</td> <td>0,9960</td> </tr> </tbody> </table>		VIII:C (%)		A405 W. Bath	A405 STAR	0	0/0	0,015	0	C/4	25/20	0,246	0,209	C/2	50/50	0,584	0,442	C	100/100	1,272	0,682	2C	200/190	2,323	0,759	Linearity: r2 =		0,9961	0,9950	VIII:C (%)		A405 W. Bath	A405 STAR	0	0/0.93	0,011	0,041	C/4	6,25/6,25	0,294	0,324	C/2	12,5/12,5	0,628	0,664	C	25/25	1,234	1,195	Linearity: r2 =		0,9990	0,9960	Water Bath: A405(200%) ≥ 1,60 ΔA405(100%-200%) ≥ 0.60 r2 ≥ 0.98 Water Bath: A405(25%) ≥ 0.80 ΔA405(12.5%-25%) ≥ 0.30 r2 ≥ 0.98	
VIII:C (%)		A405 W. Bath	A405 STAR																																																				
0	0/0	0,015	0																																																				
C/4	25/20	0,246	0,209																																																				
C/2	50/50	0,584	0,442																																																				
C	100/100	1,272	0,682																																																				
2C	200/190	2,323	0,759																																																				
Linearity: r2 =		0,9961	0,9950																																																				
VIII:C (%)		A405 W. Bath	A405 STAR																																																				
0	0/0.93	0,011	0,041																																																				
C/4	6,25/6,25	0,294	0,324																																																				
C/2	12,5/12,5	0,628	0,664																																																				
C	25/25	1,234	1,195																																																				
Linearity: r2 =		0,9990	0,9960																																																				
2. Accuracy: W.BATH <table border="1"> <thead> <tr> <th>Control</th> <th>TV* % VIII :C</th> <th>MV* % VIII :C</th> </tr> </thead> <tbody> <tr> <td>Normal Control</td> <td>94</td> <td>101</td> </tr> <tr> <td>Abnormal Control</td> <td>37</td> <td>35</td> </tr> </tbody> </table> * TV= Target Value - MV= Measured Value		Control	TV* % VIII :C	MV* % VIII :C	Normal Control	94	101	Abnormal Control	37	35	[84 - 104] [31 - 43]																																												
Control	TV* % VIII :C	MV* % VIII :C																																																					
Normal Control	94	101																																																					
Abnormal Control	37	35																																																					
3. Detection threshold Method : CS High range: < 0,5 % Low range: <0,8 %		≤ 10% ≤ 2%																																																					
4. Performances Method : CS for normal plasma and w.bath for FVIII Deficient plasma Normal plasmas: N= 15 Mean= 84 % Range: 55 to 126 % FVIII Deficient Plasma: 2,8 % VIII:C (high range) 1,2 % VIII:C (low range)		N ≥ 10 ≥ 50 % About 50-150% < 10 % < 2 %																																																					
5. Stability of restored reagents Method : WATER BATH <table border="1"> <thead> <tr> <th colspan="2">VIII:C (%)</th> <th>A405 Fresh</th> <th>A405 24H RT</th> <th>A405 72H 2-8°C</th> </tr> </thead> <tbody> <tr> <td>C/8</td> <td>13</td> <td>0,102</td> <td>0,115</td> <td>0,110</td> </tr> <tr> <td>C/4</td> <td>26</td> <td>0,229</td> <td>0,258</td> <td>0,246</td> </tr> <tr> <td>C/2</td> <td>52</td> <td>0,540</td> <td>0,592</td> <td>0,571</td> </tr> <tr> <td>C</td> <td>104</td> <td>1,177</td> <td>1,307</td> <td>1,251</td> </tr> <tr> <td>2C</td> <td>208</td> <td>2,189</td> <td>2,399</td> <td>2,393</td> </tr> <tr> <td colspan="2">Linearity: r2 =</td> <td>0,9968</td> <td>0,9969</td> <td>0,9979</td> </tr> </tbody> </table> <table border="1"> <thead> <tr> <th>Control</th> <th>MV* % VIII :C</th> <th>MV* % VIII :C</th> <th>MV* % VIII :C</th> </tr> </thead> <tbody> <tr> <td>Normal Control</td> <td>99</td> <td>101</td> <td>102</td> </tr> <tr> <td>Abnormal Control</td> <td>37</td> <td>37</td> <td>37</td> </tr> </tbody> </table> * MV= Measured Value		VIII:C (%)		A405 Fresh	A405 24H RT	A405 72H 2-8°C	C/8	13	0,102	0,115	0,110	C/4	26	0,229	0,258	0,246	C/2	52	0,540	0,592	0,571	C	104	1,177	1,307	1,251	2C	208	2,189	2,399	2,393	Linearity: r2 =		0,9968	0,9969	0,9979	Control	MV* % VIII :C	MV* % VIII :C	MV* % VIII :C	Normal Control	99	101	102	Abnormal Control	37	37	37	Δ A405 point C ≤ 15% Δ A405 point 2C ≤ 15% between fresh and 24h RT or 72H at 2-8°C [84 - 104] [31 - 43]						
VIII:C (%)		A405 Fresh	A405 24H RT	A405 72H 2-8°C																																																			
C/8	13	0,102	0,115	0,110																																																			
C/4	26	0,229	0,258	0,246																																																			
C/2	52	0,540	0,592	0,571																																																			
C	104	1,177	1,307	1,251																																																			
2C	208	2,189	2,399	2,393																																																			
Linearity: r2 =		0,9968	0,9969	0,9979																																																			
Control	MV* % VIII :C	MV* % VIII :C	MV* % VIII :C																																																				
Normal Control	99	101	102																																																				
Abnormal Control	37	37	37																																																				

Comments :



PASSED IN COMPLIANCE

Date : 20/01/2016

QC Manager : S.LECOURT