

PLASMA PAI-1 CONTROL
Référence SC011K

PLASMA PAI-1 CONTROL
SC011K

Utilisation *in vitro* exclusivement

Uniquement à usage de recherche

For *in vitro* use only

For research use only

FRANÇAIS

ENGLISH

Lot : 150703E - Exp. : 2018-05

Lot : 150703E - Exp. : 2018-05

PAI-1 Ag :

Control I lot : 141121F

Taux de PAI-1:Ag : 41.0 ng/ml

Intervalle de confiance : 35.0 à 47.0 ng/ml

Control II lot : 141121G

Taux de PAI-1:Ag : 18.0 ng/ml

Intervalle de confiance : 14.0 à 22.0 ng/ml

PAI-1 Activity :

Control I lot : 141121F

Taux de PAI-1 Activité : 9.5 ng/ml

Intervalle de confiance : 8.1 à 10.9 ng/ml

Control II lot : 141121G

Taux de PAI-1 Activité : 4.0 ng/ml

Intervalle de confiance : 3.2 à 4.8 ng/ml

PAI-1 Ag :

Control I lot : 141121F

PAI-1:Ag concentration : 41.0 ng/ml

Acceptance range : 35.0 to 47.0 ng/ml

Control II lot : 141121G

PAI-1:Ag concentration : 18.0 ng/ml

Acceptance range : 14.0 to 22.0 ng/ml

PAI-1 Activity :

Control I lot : 141121F

PAI-1 Activity concentration : 9.5 ng/ml

Acceptance range : 8.1 to 10.9 ng/ml

Control II lot : 141121G

PAI-1 Activity concentration : 4.0 ng/ml

Acceptance range : 3.2 to 4.8 ng/ml

PLASMA PAI-1 CONTROL
Référence SC011K

PLASMA PAI-1 CONTROL
SC011K

Utilisation *in vitro* exclusivement

Uniquement à usage de recherche

For *in vitro* use only

For research use only

FRANÇAIS

ENGLISH

Lot : 150703E - Exp. : 2018-05

Lot : 150703E - Exp. : 2018-05

PAI-1 Ag :

Control I lot : 141121F

Taux de PAI-1:Ag : 41.0 ng/ml

Intervalle de confiance : 35.0 à 47.0 ng/ml

Control II lot : 141121G

Taux de PAI-1:Ag : 18.0 ng/ml

Intervalle de confiance : 14.0 à 22.0 ng/ml

PAI-1 Activity :

Control I lot : 141121F

Taux de PAI-1 Activité : 9.5 ng/ml

Intervalle de confiance : 8.1 à 10.9 ng/ml

Control II lot : 141121G

Taux de PAI-1 Activité : 4.0 ng/ml

Intervalle de confiance : 3.2 à 4.8 ng/ml

PAI-1 Ag :

Control I lot : 141121F

PAI-1:Ag concentration : 41.0 ng/ml

Acceptance range : 35.0 to 47.0 ng/ml

Control II lot : 141121G

PAI-1:Ag concentration : 18.0 ng/ml

Acceptance range : 14.0 to 22.0 ng/ml

PAI-1 Activity :

Control I lot : 141121F

PAI-1 Activity concentration : 9.5 ng/ml

Acceptance range : 8.1 to 10.9 ng/ml

Control II lot : 141121G

PAI-1 Activity concentration : 4.0 ng/ml

Acceptance range : 3.2 to 4.8 ng/ml



S. LECOURT

17.07.2015