

# Prekallikrein Deficient Plasma Plasma déficient en Prékallikréine

Vial of 1 mL/flacon de 1 ml (DP310A)  
Set of 6x1mL / Coffret de 6x1 ml (DP310K)

Ref. DP310A/DP310K

STORE AT  
2-8°C

CONSERVER A  
2-8°C



Lot : F1600021

Expiration : 2018-07-03

## ANALYSIS CERTIFICATE / CERTIFICAT D'ANALYSE

ANALYTICAL DATA RESULTATS D'ANALYSE		SPECIFICATIONS SPECIFICATIONS
<b>1.</b>	<b>Prekallikrein / Prékallikréine</b>	
	FV : 87 %.	> 50%.
	APTT: 175 Sec(Reagent : CEPHEN LR-HBM)	> 50 Sec.

## CONCLUSIONS

DATE : 29/01/2016  Passed/Conforme

Refused/Non-conforme

Quality Control Manager  
*Responsable contrôle qualité*

S. LECOURT

