




Product name:	Protein L ligand leakage ELISA kit		
Production date:	2017-11		
Date of release:	2018-01-04		
Expiry date:	2018-10		
Store:	+4°C. Protein L reference at -20°C or -80°C.		
Contents:	ELISA kit for detection and quantification of ligand leakage in eluates from Protein L affinity chromatography medium containing immunoglobulin (Ig) or Ig-fragments.		
	Kit contents	Lot	Expiry date
	1 x Anti-Protein L coated microtiter plate	234508	2018-10
	130 µl HRP-conjugated anti-Protein L pAb	234705	2018-10
	200 µl Protein L reference (600 ng/ml)	231104	2019-04
	13 ml EC-Blue Enhanced™	234503	2018-10
	7 ml Stop solution, 0.5 M H ₂ SO ₄	234504	2018-10
	6 x 5 g PBS Tween tablets	234719	2018-10

Quality Control

Analysis	Specification	Result
%CV	< 5.0 %	Passed
Absorbance 60 ng/ml	> 2.1	Passed

	OD 450 nm	Concentration
Calibrator 0.94 ng/ml	0.91	-
Calibrator 1.88 ng/ml	1.36	-
Calibrator 3.75 ng/ml	1.80	-
Calibrator 7.5 ng/ml	2.21	-
Calibrator 15 ng/ml	2.46	-
Calibrator 30 ng/ml	2.70	-
Calibrator 60 ng/ml	2.81	-

Control sample A	1.81	3.7 ng/ml
Control sample B	2.03	5.2 ng/ml
Control sample C	2.48	13.3 ng/ml



The above material has met all quality specifications and has been reviewed by a quality representative

SL-Bj
 Quality Assurance, Sofia Berggren
 Medicago AB, Danmark Berga
 SE-755 08 Uppsala, Sweden

2018-01-04
 Date