

ANALYSIS CERTIFICATE

HEMOCLOT THROMBIN INHIBITORS #CK002L

Lot : F1600598

QC Release: 03/08/2016

Expiration date : 2018-12-25

Components	Qty	Exp. (months)	Int. Ref.	Lot #	Exp. date
R1 : Normal pooled plasma	3 vials	30	F161100598	F161100598	2018-12-25
R2 : Human calcium thrombin	3 vials	30	F161100598	F161100598	2019-01-04

303

ANALYSIS CERTIFICATE

HEMOCLOT THROMBIN INHIBITORS #CK002L

Lot : F1600598

QC Release: 03/08/2016

Expiration date : 2018-12-25

Analytical data	Specifications
1. <u>R1 reproducibility on basic clotting time:</u>	
Method: MC10	N ≥ 10
N : 30	≤ 3 %
CV: 1,5 %	
2. <u>R2 reproducibility on basic clotting time:</u>	
Method: MC10	N ≥ 10
N : 30	≤ 3 %
CV: 1,7 %	

803

