

Factor XIII Deficient Plasma Plasma Déficiant en Facteur XIII

Vial of 1 mL/flacon de 1 ml (DP200A)
Set of 6x1 mL/Coffret de 6x1 ml (DP200K)

Ref. DP200A / DP200K

STORE AT
2-8°C

Lot: F1600706



Expiration: 2019-01-26

CONSERVER A
2-8°C

ANALYSIS CERTIFICATE

ANALYTICAL DATA	SPECIFICATIONS
FXIII: <0.1 %	< 1%
APTT:34.1 Sec (Reagent : Cephén)	35 ± 5 Sec
FV: 80 %	> 50%

CERTIFICAT D'ANALYSE

RESULTATS D'ANALYSE	SPECIFICATIONS
FXIII: < 0.1 %	< 1%
APTT: 34.1 Sec (Réactif : Cephén)	35 ± 5 Sec
FV : 80 %	> 50%

DATE : 18/08/2016

CONCLUSIONS : Passed/Conforme

Refused/Non-conforme

Quality Control Manager
Responsable contrôle qualité

AM.VISSAC
Technical Director