

## ANALYSIS CERTIFICATE

**CALCIUM CHLORIDE 0.25M - #AR002A / AR002K**

**Lot : F1600938**

**QC release: 23/09/2016**

**Expiration date : 2019-03-08**

**Reference AR002A: 1 x 25 ml**

**Reference AR002K: 3 x 25 ml**

Components	Volume	Exp. (months)	Lot #	Exp. date
Calcium Chloride 0.25M	25 ml	30	F1600938	2019-03-08

305

# ANALYSIS CERTIFICATE

**CALCIUM CHLORIDE 0.25M - #AR002A / AR002K**

**Lot : F1600938**

**QC release : 23/09/2016**

**Expiration date : 2019-03-08**

Analytical data	Specifications
<p><b>1. <u>Batch reproductibility</u></b></p> <p style="text-align: right;">N : 3</p> <p style="text-align: right;">CV(CT) 0,3 %</p> <p style="text-align: right;">Volume 26 ml</p> <p style="text-align: left;">N = 5</p>	<p><i>Indicative</i> ≥ 25 ml</p>
<p><b>2. <u>Aspect</u></b> Clear, transparent no sign of contamination</p>	<p>Clear, transparent, no sign of contamination</p>
<p><b>3. <u>aPTT on citrated human plasmas</u></b></p> <p style="text-align: right;">Instrument : MC10</p> <p style="text-align: right;">Cephalin : HBM- Cephen LR</p> <p>Normal plasmas</p> <p style="text-align: right;">N : 5</p> <p style="text-align: right;">Mean CT : 33 Sec.</p> <p style="text-align: right;">ΔCT (control sample) &lt;1 Sec.</p> <p>Pathological plasmas</p> <p style="text-align: right;">N: 5</p> <p style="text-align: right;">Mean CT : 97 Sec.</p> <p style="text-align: right;">ΔCT (control sample) 2 %</p>	<p>N ≥ 5 35 ± 5 Sec ≤ 2 sec</p> <p>N ≥ 5 &gt; 40 Sec ≤ 5%</p>
<p><b>4. <u>FVIII:C clotting assay</u></b></p> <p style="text-align: right;">Instrument: MC10</p> <p style="text-align: right;">CT ( 100 %): 57 Sec.</p> <p style="text-align: right;">ΔCT (0-100%) 80 Sec.</p> <p style="text-align: right;">ΔCT (control sample) &lt;1 %</p>	<p>CT (100%)&lt; 75 sec ΔCT(0-100%) ≥ 40 sec ≤ 5%</p>
<p><b>Comments :</b></p>	<p><input checked="" type="checkbox"/> <b>PASSED IN COMPLIANCE</b></p>

Date : 23/092016

QC Manager :

S.LECOURT

