

## ANALYSIS CERTIFICATE

**BIOPHEN™ Rivaroxaban Calibrator Low - #226001**

**Lot : F1601010**

**QC release: 2016-11-22**

**Expiration date : 2019-04-02**

<b>Components</b>	<b>Volume</b>	<b>Exp. (months)</b>	<b>Lot #</b>	<b>Exp. date</b>
CAL I : Calibrator 1	4 vials	30	F161101010	2019-04-02
CAL II : Calibrator 2	4 vials	30	F161101010	2019-04-02
CAL III : Calibrator 3	4 vials	30	F161101010	2019-04-02



## ANALYSIS CERTIFICATE

**BIOPHEN™ Rivaroxaban Calibrator Low - #226001**

**Lot : F1601010**

**QC release: 2016-11-22**

**Expiration date : 2019-04-02**

Analytical data	Specifications
<b>1. <u>Within lot reproducibility</u></b>	
N= 10                      CV:            0,5 %	CV (OD) ≤ 2%
N= 10                      CV:            0,4 %	CV (OD) ≤ 2%
N= 10                      CV:            0,8 %	CV (OD) ≤ 2%

2. <u>Concentration [C] and Standard Deviation (SD)</u>			
<b>Controls</b>	<b>[C] ng/mL</b>	<b>SD</b>	
CAL I	0	1,74	CAL1: < 20 ng/ml
CAL II	51	2,94	CAL2: 40-60 ng/mL
CAL III	109	4,39	CAL3: 80-120 ng/mL

3. <u>Aspect</u>		
<input checked="" type="checkbox"/>	Slightly opalescent to clear	a) Slightly opalescent to clear
<input checked="" type="checkbox"/>	No coagulum	b) No coagulum
<input checked="" type="checkbox"/>	Stable solution	c) Stable solution

4. <u>Stability of reconstituted reagents</u>					
		Fresh	48h	7 days	
		/	RT	2-8°C	
CAL I	ng/mL	0	0	0	<p style="text-align: center;"><b><u>48 hours at RT:</u></b></p> <p style="text-align: center;">Δ [C] ≤ 15 ng/ml</p> <p style="text-align: center;"><b><u>7 days at 2-8°C:</u></b></p> <p style="text-align: center;">Δ [C] ≤ 15 ng/ml</p>
	Δ [C]	NA	0	0	
CAL II	ng/mL	51	51	52	
	Δ [C]	NA	0	1	
CAL III	ng/mL	111	111	110	
	Δ [C]	NA	0	1	

95

## ANALYSIS CERTIFICATE

**BIOPHEN™ Rivaroxaban Calibrator Low - #226001**

**Lot : F1601010**

**QC release: 2016-11-22**

**Expiratin date : 2019-04-02**

Analytical data	Specifications												
<p><b>5. Calibration curve</b></p> <p style="text-align: center;"><input type="checkbox"/> Manual method      <input checked="" type="checkbox"/> STAR</p> <p style="text-align: center;"><b>BIOPHEN DiXal                      Lot F1600443</b></p> <table border="1" style="margin-left: auto; margin-right: auto; border-collapse: collapse;"> <thead> <tr> <th style="width: 10%;"></th> <th style="width: 20%;">ng/ml</th> <th style="width: 20%;">A<sub>405</sub></th> </tr> </thead> <tbody> <tr> <td style="text-align: center;">CAL I</td> <td style="text-align: center;">0</td> <td style="text-align: center;">2,313</td> </tr> <tr> <td style="text-align: center;">CAL II</td> <td style="text-align: center;">51</td> <td style="text-align: center;">1,436</td> </tr> <tr> <td style="text-align: center;">CAL III</td> <td style="text-align: center;">109</td> <td style="text-align: center;">0,753</td> </tr> </tbody> </table>		ng/ml	A <sub>405</sub>	CAL I	0	2,313	CAL II	51	1,436	CAL III	109	0,753	
	ng/ml	A <sub>405</sub>											
CAL I	0	2,313											
CAL II	51	1,436											
CAL III	109	0,753											

<p><b>6. Linearity</b></p> <p style="text-align: center;"><b>R<sup>2</sup>                      0,998</b></p>	<p><b>R<sup>2</sup> ≥ 0.98</b></p>
---	------------------------------------

<p><b>7. Accuracy</b></p> <p style="text-align: center;"><input type="checkbox"/> Manual method      <input checked="" type="checkbox"/> STAR</p> <table border="1" style="margin-left: auto; margin-right: auto; border-collapse: collapse;"> <thead> <tr> <th colspan="3" style="text-align: center;">CONTROLS</th> <th style="text-align: center;">TV*</th> <th style="text-align: center;">MV*</th> </tr> </thead> <tbody> <tr> <td style="text-align: center;">C I</td> <td style="text-align: center;">Lot</td> <td style="text-align: center;">50502-1</td> <td style="text-align: center;">25</td> <td style="text-align: center;">24</td> </tr> <tr> <td style="text-align: center;">C II</td> <td style="text-align: center;">Lot</td> <td style="text-align: center;">50502-2</td> <td style="text-align: center;">76</td> <td style="text-align: center;">70</td> </tr> </tbody> </table> <p style="text-align: center; font-size: small;">*TV: Target Value      *MV: Measured Value</p>	CONTROLS			TV*	MV*	C I	Lot	50502-1	25	24	C II	Lot	50502-2	76	70	<p>MV* within the acceptance range</p> <p style="text-align: center;">[ 15 - 35 ] [ 61 - 91 ]</p>
CONTROLS			TV*	MV*												
C I	Lot	50502-1	25	24												
C II	Lot	50502-2	76	70												

<p>Comments :</p>	<p><input checked="" type="checkbox"/> <b>PASSED IN COMPLIANCE</b></p>
-------------------	--

**Date : 2016-11-22**

**QC Manager : S.LECOURT**



**BIOPHEN® RIVAROXABAN CALIBRATOR LOW**  
Reference 226001

Pour diagnostic *in vitro* exclusivement.

FRANÇAIS

Lot : F1601010 - Exp. : 2019-04-02

Concentration [C] en Rivaroxaban dans les calibrateurs

Cal 1 Lot : F161101010  
[C] : 0 ng/mL

Cal 2 Lot : F161101010  
[C] : 51 ng/mL

Cal 3 Lot : F161101010  
[C] : 109 ng/mL

**BIOPHEN® RIVAROXABAN CALIBRATOR LOW**  
Reference 226001

For *in vitro* diagnostic use only

ENGLISH

Lot : F1601010 - Exp. : 2019-04-02

Rivaroxaban concentration [C] in the calibrators

Cal 1 Lot : F161101010  
[C] : 0 ng/mL

Cal 2 Lot : F161101010  
[C] : 51 ng/mL

Cal 3 Lot : F161101010  
[C] : 109 ng/mL

**BIOPHEN® RIVAROXABAN CALIBRATOR LOW**  
Reference 226001

Pour diagnostic *in vitro* exclusivement.

FRANÇAIS

Lot : F1601010 - Exp. : 2019-04-02

Concentration [C] en Rivaroxaban dans les calibrateurs

Cal 1 Lot : F161101010  
[C] : 0 ng/mL

Cal 2 Lot : F161101010  
[C] : 51 ng/mL

Cal 3 Lot : F161101010  
[C] : 109 ng/mL

Quality Control Approval  
23 NOV. 2016  
S. Lecourt

**BIOPHEN® RIVAROXABAN CALIBRATOR LOW**  
Reference 226001

For *in vitro* diagnostic use only

ENGLISH

Lot : F1601010 - Exp. : 2019-04-02

Rivaroxaban concentration [C] in the calibrators

Cal 1 Lot : F161101010  
[C] : 0 ng/mL

Cal 2 Lot : F161101010  
[C] : 51 ng/mL

Cal 3 Lot : F161101010  
[C] : 109 ng/mL